



OFFICE OF THE PRINCIPAL
DIAMOND HARBOUR GOVERNMENT MEDICAL COLLEGE
DEPARTMENT OF HEALTH AND FAMILY WELFARE
GOVERNMENT OF WEST BENGAL

Ph: 03174-255-633

E-mail: diaharmc@gmail.com

MEMO No. DHGMC/2021/1558

Date: 26/10/2021

Corrigendum Notice No:1

(Ref: NIT No: 22, e-tender Notice No: DHGMC/2021/1364 dt; 14/09/2021)

In reference to the above said matter, the below mentioned corrigendum are hereby included in the bid document (DHGMC/2021/1361 dt; 14/09/2021).

As per unanimous decision the pre bid meeting dated 28/09/2021 in presence of interested bidders and TIA.

• **Addendum Specification of HD Laparoscopy Set:**

1. The last date for submission of the above said e-tender was 30/10/2021, the TIA, DHGMCH, Diamond Harbour, has decided to extend the date for submission of e-tender up to 16/11/2021, 11:00 PM & Technical bid opening date- 19/11/2021, after 11:00 AM.
2. The tender committee, DHGMCH and TIA, DHGMCH are hereby requested to submit the quoted rate in the BOQ, excluding Ultrasonic Shear (Page No 15 in the Bid Document). The corrigendum specification attached herewith

Principal

Diamond Harbour Govt. Medical College.
Diamond Harbour, South 24 Pgs.

Dated:- 26/10/2021

Memo No: DHGMC/2021/1558/1/(8)

Copy forwarded for information and necessary action to the:-

1. The Director of Medical Education & Ex-Officio Secretary to the Govt. of West Bengal, Dept. of Health and Family Welfare, Swasthya Bhawan, GN-29, Sector-V, Salt Lake City, Kolkata-700 091.
2. The District Magistrate, South 24 pgs .
3. The Medical Superintendent-cum-Vice Principal, DHGMCH, South 24 pgs
4. The Sub Divisional Officer , Diamond Harbor, South 24 Parganas
5. The HOD, Dept. of General Surgery, DHGMCH, Diamond Harbour.
6. The Chairman, Diamond Harbour Municipality, South 24 Parganas
7. The Accounts Officer,, DHGMC, S 24 Pgs
8. All members of the Tender & Purchase Committee, DHGMCH, S 24Pgs
9. IT Cell, Department of Health & F.W. Swasthya Bhawan- with a request to upload this notice in the Health Departmental website.
10. The District Information Officer, NIC, South 24 pgs Collectorate Bldg., Alipore -wih a request upload this notice in the Departmental website.
11. Guard File.

Principal

Diamond Harbour Govt. Medical College.
Diamond Harbour, South 24 Pgs.

Desired Specifications for HD Laparoscopy Set with accessories

S. No	Desired Specifications	Bidder's offer with variations
A	Endo-vision camera & CCD unit / CMOS	
1	High Definition Three Chip Camera System	
2	220.V, AC Camera console with universal coupler	
3	Pure Digital signal for high definition video (at least 1920 x 1080 native resolution)	
4	Resolution-2000 horizontal lines with multiple specialty settings	
5	High Noise Immunity & Enhanced Light Sensitivity	
6	Progressive scan technology (CMOS High Definition technology)	
7	Brightness Control on both console & camera head	
8	Aperture Control on console.	
9	Inbuilt 16 step digital Image Enhancer in console	
10	Digital zoom & white balance on camera head	
11	Integrated Gain/shutter/Enhancement with brightness control	
12	Two peripheral control on camera head	
13	Should have 2 DVI, 2 SVHS, 1 RGB output and 1 composite output	
B	Light Source	
1	220 V, 250 Watt (minimum) LED lamps with least 40,000 hrs of lamp life	
2	Should preferably have digital lamp life counter	
3	Light intensity adjustment: continuously adjustable from 0 to 100% manually. Can have Automatic adjustment option additionally	
4	Must have Standby mode	
5	Universal Jaw Assembly to adapt Fibre Optic Cable of any make, without adapter.	
C	Fibre Optic Cable	
1	Cable diameter should be 5.5 mm. Length should be 7.5 feet	
2	Should be snap fit and should have universal coupling system at the patient end to match all varieties of telescopes	
D	Monitor	
1	Medical Grade, flat pane 26" LED monitor	
2	Should have compatible resolution for best output of HD Video signals	

Sub-brother Ray
19/7/17

Don
19.7.17

Slant
19.7.17

Principal
19.07.17
Prof. S.K. Mukherjee
Principal
CMSDH, Kolkata-68

3	Should have DVI, SVHS, RGB and Composite ports	
E	Insufflator	
1	Microprocessor controlled high flow (up to 40 l/min)	
2	Built in heating system	
3	Soft Approach Pressure control for safe recovery of abdominal pressure	
4	LCD based central display monitor	
5	Both Audio & Visual warning signals	
6	Should have detection system for internal leakage	
F	Telescope (HD)	
1	Fully autoclavable	
2	Working length – 30 cm	
3	Universal adapter for light cables of other make	
4	Yellow Glass index for optimum evenness of focus & contrast	
5	Should have tamper proof, non reflective SS Housing or optical fibres and rod lens system	
6	Sizes – 0°, 5 and 10 mm; 30°, 5 mm and 10 mm	
G	Suction irrigation apparatus	
1	Capable of pouring high speed water jet (at least 2 l/min) through 5 mm suction cannula	
2	Should have option to control the water flow as well as rate of suction	

List of Hand Instruments and other Accessories which need to be provided along with each set –

Laparoscopic hand instruments (reusable) with 310mm working length, take apart locking / unlocking mechanism, rotatable with interchangeable handle and insert and monopolar diathermy attachment (Except trocar-cannula and Verees needle and suction-irrigation cannula).

1. Verees needle 12 cm length - 2 pcs
2. Verees needle 15 cm length - 2 pcs
3. Silicon-PVC tubing for CO₂ flow – 4 pcs
4. 11 mm and 5.5 mm Trocar sleeves – 4 pcs each
5. 10 mm and 5 mm Trocars (Pyramidal tip) – 4 pcs each
6. 11-5 mm reducer sleeve type and flip on type – 2 pcs each
7. Silicon-Rubber washers, 5 mm – 100 pcs
8. Silicon-Rubber washers, 10 mm – 50 pcs
9. Maryland dissector, 5mm – 3 pcs
10. Atraumatic graspers, 5mm, stout – 2 pcs
11. Toothed graspers, 5 mm – 2 pcs
12. Metzenbaum scissors – 3 pcs
13. Bowel graspers, 5 mm – 2 pcs
14. Hook electrode, L type, 5 mm – 3 pcs
15. Suction-Irrigation nozzle, 5 mm with Toggle/Trumpet type valve – 2 pcs
16. Suction-Irrigation nozzle, 10 mm with Toggle/Trumpet type valve – 1 pc
17. Stone retrieving forceps, 10 mm – 1 pc

Debabrati Ray
19/7/12

Don
19.7.12

Shanta
19.7.12

Shanta
19.7.12

18. Laparoscopic biopsy forceps, 5 mm – 1 pc
19. Clip Applicator, 10 mm, for LT300 & LT400 clips – 2 pcs
20. Laparoscopic Needle driver and grasper, 5 mm – 2 pairs
21. Gall bladder Extraction Forceps, 10 mm – 2 pcs
22. Fan retractors 5 mm – 2 pcs
23. Port closure needle – 2 pcs
24. Hassan Canula, 10 mm – 1 pc
25. Diathermy cable (cautery lead) – 5 pcs
26. SS sterilization tray with cover – 3 pcs
27. Lap endotrainer – 1 pc
28. 12 mm Trocar with sheeth – 2 pcs (wherever deemed necessary)
29. Silicon-Rubber washers, 12 mm – 20 pcs (wherever deemed necessary)
30. Hemostatic clip applicator – 1pc (wherever deemed necessary)
31. Bipolar, 5 mm Maryland Dissector – 1 pc
32. Bipolar grasper, 5 mm – 1 pc
33. Bipolar Diathermy cable – 3 pcs
34. Monopolar diathermy electrode, 5 mm, ball type – 1 pc
35. Monopolar diathermy electrode, 5 mm, spatula type – 1 pc

European CE and US FDA certification needed for all the components (HD Laparoscopy set and Accessories)

M
26/10/2021

Principal
Diamond Harbour Government Medical College & Hospital
Diamond Harbour, South 24 Parganas

Madhumita Mukherjee

PROF. & HOD

DEPT.....

Diamond Harbour Govt. Medical College & Hospital
South 24 Parganas

26.10.2021

Debabrudi Roy
19/2/17

Tan
19.2.17

Stent
19.7.12

Prof. S.K. Mukherjee
Principal
CMSDH, Kolkata-58
19.07.12